



Republic of the Philippines
Department of Health
REGION 1 MEDICAL CENTER
Arellano Street, Dagupan City
BAC Office
Tel No.: (075) 523-18-65
Telefax.: (075) 523-41-03
r1mcsbac2012@gmail.com

NOTICE OF AWARD

January 26, 2018

EDGARDO O. BOLOMPO JR.
Owner / Manager
DREI ENTERPRISES
#06 Purvenir St. Central East
Buang La Union

Sir:

We are informing you that your Bid dated **January 17, 2018** for the **PROCUREMENT OF VARIOUS HOSPITAL EQUIPMENT, under Invitation to Bid (IB) No. 1, s, 2018, Bid Reference No. 01-18 is AWARDED** to you for an amount equivalent to One Peso (Php1.00) as consideration for the option granted to the Region 1 Medical Center to buy the items (listed as **Annex "A"**) pursuant to the Revised Guidelines on the Use of Ordering Agreement, in accordance to the Revised IRR of RA 9184.

You are therefore required to provide within ten (10) calendar days the performance security in an amount equal to the percentage of the total contract price in accordance with ITB Clause No. 32 of Section II – Instruction to Bidders. Failure to provide the performance security shall constitute sufficient ground for cancellation of the award and forfeiture of the bid security.

Approved By:


JOSEPH ROLAND O. MEJIA, MD, FPSMS, MBAH, MPA, DHSM, DCDM, CESE,
Medical Center Chief II

Conforme:


(Name of Representative of Bidder)


(Authorized Signature)

January 29 2018
(Date)

Performance Security:

Form of PS: G(13)-A No. 004467
Performance Bond

Amount: ₱ 4,620,000.00

Date: January 30 2018



Republic of the Philippines
Department of Health
REGION 1 MEDICAL CENTER
Arellano Street, Dagupan City
BAC Office
Tel No.: (075) 523-18-65
Telefax.: (075) 523-41-03
r1mcsbac2012@gmail.com

ANNEX A

Procurement of Various Hospital Equipment

LINE / PARTIAL

Bid Ref. 01-2018

Item / Description	Qty	Unit	Unit Cost	Total	Remarks
LAPAROSCOPY TOWER SYSTEM	1	Unit	15,400,000.00	15,400,000.00	
Please see attached Specification					
Total Bid Price	Php 15,400,000.00				
Total Bid Price in Words	Fifteen Million Four Hundred Thousand Pesos				
Name & Address of the Bidder	DREI ENTERPRISES #06 Purvenir St. Central East Buang La Union				

LAPAROSCOPY TOWER SYSTEM

Imaging system

- High Definition LCD Monitor 32" medical grade, 16:10 aspect ratio resolution 1920x1200 pixels, with Monitor Stand
- High Definition Camera Control Unit with USB thumb drive socket, 1920x1200 pixels, compatible with leading brands of EGD Scope, ERCP scope, colonoscopy, and bronchoscope
- Xenon Light Source 300W, 500 hours with natural light filter option for early cancer detection feature 300W compatible with leading brands of EGD Scope, ERCP scope, colonoscopy, and bronchoscope
- High Definition Camera Head, 3 CCD Chips autoclavable eyepiece type with 3 programmable remote switches, 2 zoom buttons, 2 focus buttons, integrated with coupler
- Light Guide Cable, M plug type, 3 meters
- High Flow Insufflation Unit, 45 liters per minute high flow insufflator high, medium and low flow rate control with insufflation tube high pressure hose with smoke evacuation feature Pump 220-240V for irrigation in laparoscopy use and arthroscopy with reusable tubing set
- High Definition telescope 10mm 30 degree DOV 310mm working length
- High Definition telescope 10mm 0degree DOV 310mm working length
- Includes the local trolley DVD recorder and AVR

- Instruments for Lap Cholecystectomy, Lap AP, lap OB with inclusion of clip applicator for medium and large Trocar Set
- Bipolar high frequency cable 3.5m length
- Claw Forceps 330mm WL, 28mm jaw, 5mm diameter
- Claw Forceps 330mm, 10x330mm, 45mm jaw length
- Clinching Forceps 5mmx330mm, 26mm jaw
- Handle set for suction irrigation tube 5mm with suction irrigation lever for suction/irrigation tube
- Hook electrode with suction channel 5x330mm, monopolar
- Johann forceps 5mmx330mm, 16.5mm jaw bipolar
- Long Johann grasping forceps, 5 x 330 mm, 40mm jaw length
- Manhes forceps 330mm, 14mm jaw, 5mm diameter
- Maryland forceps 5mmx330mm, 14.7 mm jaw bipolar
- Metzenbaum scissors curved 5 x 330mm, 19mm jaw length
- Monopolar cautery Cable 3.5m length
- Monopolar needle electrode 5x330mm
- Needle holder left curved (for right handed users) 5x330mm length
- Short Maryland forceps 5x330mm, 21mm jaw length
- Suction irrigation tube 5.3mm x 360mm with monopolar button tip
- Standard requirement (Current and Valid Certificate of Manufacturer's compliance with ISO and/or CE Certificate, Original Brochure, User's Manual/Brochure in English Language, Operations Manual 2 copies, Service Manual 2 copies
- Training requirements (Training at least two (2) end-users),
- After Sales (Bidder's certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the warranty period)
- Warranty certificate (Minimum of one (1) year for parts and service, upon delivery, inspection and acceptance)

w/ spare bulbs

lifetime service warranty

