



NOTICE OF AWARD

April 6, 2021

EDGARDO BOLOMPO JR.

Owner

DRIE ENTERPRISES

Bauang La Union

Sir:

This is to inform you that the Equipment Repair, O for ^{Medical Equipment} per BAC Resolution No. AMP-NP-022A s, (3) 2021 through the use of Alternative Mode of Procurement, specifically **Small Value Procurement** is hereby awarded to you at a Contract Price equivalent to **Ninety- Nine Thousand Eight Hundred Pesos (P 99,800.00)**.

ITEM NO.	ITEM & DESCRIPTION	QTY.	UNIT	UNIT PRICE	TOTAL PRICE
1.	Equipment Repair, O Are/Section: Pedia ward Equipment: ECG machine Brand: Mortara Model: ELI 230 Serial Number: 112290126407 Remarks: Complaint: no printing output Findings: Defective ECG Cable and Printer Roller Work Performed: General Check up and repair of unit Recommendation: Need to replace ECG cable Status: ECG machine is Functional Area/Section: Recovery Room Equipment: ECG Machine Brand: Mortara Serial number: 115130251791 Remarks: Complaint: ECG output Interference Findings: Defective ECG patient Cable Action Taken: Checked up the unit with simulator, machine had in good condition Recommendation: purchase a new ECG Patient Cable	2	Lot	49,900.00	99,800.00
GRAND TOTAL					99,800.00

You are therefore required to deliver the above items within the period stated in the Request for Quotation

You have submitted, Failure to comply shall constitute a sufficient ground for **cancellation of this award**.

Approved By:

JOSEPH ROLAND O. MEJIA, MD, FPSMS, MBAH, MPA, DHSM, DCDM, CESE
 Medical Center Chief II

Conforme:

(Name of Representative of Bidder)

(Authorized Signature)

(Date)