



NOTICE OF AWARD

ROMELYN JOYCE M. BERNARDINO

Authorized Representative

ALCON LABORATORIES (PHIL), INC.

No. 801 8/F Tower Rockwell Center Meralco

Compound Ortigas Avenue City Pasig City Metro Manila

Sir/Mam:

This is to inform you that the Equipment Repair per **BAC Resolution No. AMP-NP-170A s, (10) 2020** through the use of Alternative Mode of Procurement, specifically **Small Value Procurement** is hereby awarded to you at a Contract Price equivalent to **One Hundred Thirty-Six Thousand Seven Hundred Seventy Pesos (P 136,770.00)**.

ITEM NO.	ITEM & DESCRIPTION	QTY.	UNIT	UNIT PRICE	TOTAL PRICE
1	EQUIPMENT REPAIR, Q-5 Medical Equipment Corrective Form Area/Section: OPD-OPHTHA Equipment; Phacoemulsification Brand: Constillation Visions Model: ALCON Serial Number : 1403179701X Date Acquired : 2015 Control Number : MECF100720-A Date/Time: 10-07-20 @ 9AM Remarks: Complaint: for check up Findings: Functionality check-up of Machine found " System Error 3420" include infusion pressure transducer offset error. Infusion in 30 MMHG back-up mode. Irrigation and extraction are disabled. Action Taken: Contacted the supplier of the machine and found replacement of the Assy,PCB,Fluidics Infusion RHS; Machine Status: For Repair (See Attached Quotation/Service of Supplier).	1	Lot	136,770.00	136,770.00
GRAND TOTAL					₱ 136,770.00

You are therefore required to deliver the above items within the period stated in the Request for Quotation

You have submitted, Failure to comply shall constitute a sufficient ground for **cancellation of this award.**

Approved By:

JOSEPH ROLAND O. MEJIA, MD, FPSMS, MBAH, MPA, DHSM, DCDM, CESE
Medical Center Chief II

Conforme:

Romelyn Joyce Bernardino
(Name of Representative of Bidder)

[Signature]
(Authorized Signature)

(Date)