



Republic of the Philippines  
Department of Health  
**REGION 1 MEDICAL CENTER**  
Arellano Street, Dagupan City  
**BAC Office**  
Tel No.: (075) 523-18-65  
Telefax: (075) 523-41-03  
[r1mcsbac2012@gmail.com](mailto:r1mcsbac2012@gmail.com)

## NOTICE OF AWARD

March 10, 2022

**EDGARDO BOLOMPO JR.**

Owner

**DRIE ENTERPRISES**

Bauang La Union

Sir:

This is to inform you that the Equipment Repair per **BAC Resolution No. AMP-NP-021A s, (2) 2022** through the use of Alternative Mode of Procurement, specifically **Small Value Procurement** is hereby awarded to you at a Contract Price equivalent to **Ninety-Nine Thousand Nine Hundred Sixty Thousand Pesos (P 99,960.00)**.

ITEM NO.	ITEM & DESCRIPTION	QTY.	UNIT	UNIT PRICE	TOTAL PRICE
5.	EQUIPMWNT REPAIR, 0-2 (Estimated Cost Range from 45,001.00 to 50,000.00) ( 9303) Area/Section: Emergency Room; Equipment: Defibrillator; Brand: MINDRAY; Model: D6; Serial Number: DG-2A007141; Control Number: MECF1011022-E; Complaint: No-Back-up Power, Battery aged alarm; Action Take: Physical inspection and User test calibration; Findings: Defective back up battery and ECG cable with extension cable; Machine Status: For Replacement of 14.8v/300mah Rechargeable Li-Ion Battery with charge indicator and for replacement of ECG Cable and ECG Cable.	1	lot	49,980.00	49,980.00
6.	EQUIPMENT REPAIR, 0-2 (Estimated Cost Range from 45,001.00 to 50,000.00) ( 9303) Area/Section: ICU-CCU; Equipment: defibrillator; Brand; MINDRAY; Model; D3; Serial Number: MECF011022-C; Complaint: No back up power, battery aged alarm; Findings: Defective back up battery; Action Taken: Standby; For Replacement of 6 pin 14.8V/3000MAH Li-Ion Battery with charge indicator and for replacement of ECG cable with ECG Extension Cable	1	lot	49,980.00	49,980.00
<b>GRAND TOTAL</b>				<b>99,960.00</b>	


You are therefore required to deliver the above items within the period stated in the Request for Quotation You have submitted, Failure to comply shall constitute a sufficient ground for **cancellation of this award**.

Approved By:

  
**JOSEPH ROLAND O. MEJIA, MD, FPSMS, MBAH, MPA, DHSM, DCDM, CESE**  
Medical Center Chief II

Conforme:

  
(Name of Representative of Bidder)

  
(Authorized Signature)

3-10-2022

(Date)

R1MC-MCC-BAC-FRM-021

REV.0/25-JULY-2013