



Republic of the Philippines  
Department of Health  
**REGION 1 MEDICAL CENTER**  
Arellano Street, Dagupan City  
**BAC Office**  
Tel No.: (075) 523-18-65  
Telefax: (075) 523-41-03  
[r1mcsbac2012@gmail.com](mailto:r1mcsbac2012@gmail.com)

## NOTICE OF AWARD

March 16, 2022

**EDGARDO BOLOMPO JR.**  
Owner  
**DRIE ENTERPRISES**  
Bauang La Union

Sir:

This is to inform you that the Equipment Repair, O-2 per **BAC Resolution No. AMP-NP-024A s, (3) 2022** through the use of Alternative Mode of Procurement, specifically **Small Value Procurement** is hereby awarded to you at a Contract Price equivalent to **Forty-Nine Thousand Five Hundred Pesos (P 49,500.00)**.


ITEM NO.	ITEM & DESCRIPTION	QTY.	UNIT	UNIT PRICE	TOTAL PRICE
2.	EQUIPMENT REPAIR, O-2 (Estimated Cost Range from 45,001.00 to 50,000.00)(9303) Area/Section: Operating Room, Equipment: Patient Monitoring, Brand: Advance Medical, Model: AM-2000E, Serial Number: 2015059203, Complaint: Not Working; for check up, Findings: Defective SPO2 Sensor; NIBP Cuff; NIBP Hose; ECG Cable; CMOS Battery and NIBP Module, Action Taken: General Check up and functionality Test, Recommendation: For Replacement of Defective Accessories/parts. Machine Status: Stand By: For Repair	1	Lot	49,500.00	49,500.00
<b>GRAND TOTAL</b>					<b>49,500.00</b>

You are therefore required to deliver the above items within the period stated in the Request for Quotation  
You have submitted, Failure to comply shall constitute a sufficient ground for **cancellation of this award**.

Approved By:

  
**JOSEPH ROLAND O. MEJIA, MD, FPSMS, MBAH, MPA, DHSM, DCDM, CESE**  
Medical Center Chief II

Conforme:

  
(Name of Representative of Bidder)

  
(Authorized Signature)

3/16/2022  
(Date)

R1MC-MCC-BAC-FRM-021

REV.0/25-JULY-2013