



Republic of the Philippines
Department of Health
REGION 1 MEDICAL CENTER
Arellano Street, Dagupan City
BAC Office
Tel No.: (075) 523-18-65
Telefax: (075) 523-41-03
r1mcsbac2012@gmail.com

BAC COPY

NOTICE OF AWARD

June 3, 2022

EDGARDO BOLOMPO JR.

Owner

DRIE ENTERPRISES

Bauang La Union

Sir:


This is to inform you that the Equipment Repair of X-ray per **BAC Resolution No. AMP-NP-080A s, (5) 2022** through the use of Alternative Mode of Procurement, specifically **Small Value Procurement** is hereby awarded to you at a Contract Price equivalent to **One Hundred Forty-Nine thousand Fifty Pesos (P 149,050.00)**.

ITEM NO.	ITEM & DESCRIPTION	QTY.	UNIT	UNIT PRICE	TOTAL PRICE
1.	Equipment Repair, Q-2 (Estimated Cost Range From 140,001.00 to 150,000.00) (9321) Medical Equipment Corrective Form Area/Section:Radiology Equipment:X-ray Brand: Fuji Film Model: FDR Smart X Serial Number:DXB1840103 Control Number: MECF011422-C Date/Time: 01-14-22 @ 9:30 AM Remarks: Comlaint: For check-up Findings: Contrl Head Buttons/ Panel cannot rotate up and down, left and right and table cannot move/slide. Action Taken: Physical inspection, general check-up and functionality test. Recommendation: According to the authorize technician is subject for replacement of PMB (see attached technician service report) Machine Status: Machine is not fully operational/working	1	Lot	149,050.00	149,050.00
GRAND TOTAL					149,050.00


You are therefore required to deliver the above items within the period stated in the Request for Quotation

You have submitted, Failure to comply shall constitute a sufficient ground for **cancellation of this award**.

Approved By:


JOSEPH ROLAND O. MEJIA, MD, FPSMS, MBAH, MPA, DHSM, DCDM, CESE
Medical Center Chief II

Conforme:


(Name of Representative of Bidder)


(Authorized Signature)

(Date)

R1MC-MCC-BAC-FRM-021

REV.0/25-JULY-2013